



All dimensions are in mm; tolerances according to ISO 2768 m-H

**Interface**

According to

IEC 61169-4, EN 122190, DIN 47223

**Contents and Documentation**

This kit is delivered with

- **Standard Definitions Card**  
Printed Standard Definitions that can be used on nearly all Vector Network Analyzers
- **Certificate of Testing**
- **Lanyard**
- **Hard Shell Case**
- **Protection Caps**
- **User Manual**

**Material and plating**

**Connector parts**

Center conductor  
Outer conductor  
Body  
Dielectric  
Substrate

**Material**

CuBe  
Brass  
Brass  
PP  
Al<sub>2</sub>O<sub>3</sub>

**Plating**

Gold, min. 1.27 µm, over nickel  
Flash white bronze over silver(e.g. Optargen®)  
powder coated

**Electrical data**

Frequency range	DC to 4 GHz
<b>Open</b>	
Return loss	≤ 0.15 dB, DC to 4 GHz
Error from nominal phase <sup>1</sup>	≤ 3.0°, DC to 4 GHz
<b>Short</b>	
Return loss	≤ 0.15 dB, DC to 4 GHz
Error from nominal phase <sup>2</sup>	≤ 3.0°, DC to 4 GHz
<b>Load</b>	
Return loss	≥ 40 dB, DC to 2.5 GHz
	≥ 38 dB, 2.5 GHz to 4 GHz
DC-Resistance	50 Ω ± 0.5 Ω
Power handling (at 25 °C, sea level)	≤ 1.0 W, derate by 0.01 W/K

<sup>1</sup> The nominal phase is defined by the Offset Delay, the Offset Loss and the Fringing Capacitances

<sup>2</sup> The nominal phase is defined by the Offset Delay, the Offset Loss and the Short Inductance

**Mechanical data**

Mating cycles	≥ 500
Maximum torque	30 Nm
Recommended torque	2.26 Nm
Gauge	1.77 mm to 2.07 mm

**General standard definitions**

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

**Open**

Offset Z <sub>o</sub> / Impedance / Z <sub>o</sub>	50 Ω
Offset Delay	87.394 ps
Length (electrical) / Offset Length	26.20 mm
Offset Loss	0.50 GΩ/s
Loss	0.0076 dB/√GHz
Fringing Capacitances	C <sub>0</sub> = 177.000 x 10 <sup>-15</sup> F / 177.000 fF
	C <sub>1</sub> = 7200.00 x 10 <sup>-27</sup> F/Hz / 7.20000 fF /GHz
	C <sub>2</sub> = -3300.00 x 10 <sup>-36</sup> F/Hz <sup>2</sup> / -3.30000 fF /GHz <sup>2</sup>
	C <sub>3</sub> = 386.000 x 10 <sup>-45</sup> F/Hz <sup>3</sup> / 0.38600 fF /GHz <sup>3</sup>

**Short**

Offset $Z_o$ / Impedance / $Z_o$	50 $\Omega$		
Offset Delay	96.734 ps		
Length (electrical) / Offset Length	29.00 mm		
Offset Loss	0.50 G $\Omega$ /s		
Loss	0.0084 dB/ $\sqrt{\text{GHz}}$		
Short Inductance	$L_0 = 0.0000 \times 10^{-12}$ H	/	0.0000 pH
	$L_1 = 0.0000 \times 10^{-24}$ H/Hz	/	0.0000 pH/GHz
	$L_2 = 0.0000 \times 10^{-33}$ H/Hz <sup>2</sup>	/	0.0000 pH/GHz <sup>2</sup>
	$L_3 = 0.0000 \times 10^{-42}$ H/Hz <sup>3</sup>	/	0.0000 pH/GHz <sup>3</sup>

**Load**

Offset $Z_o$ / Impedance / $Z_o$	50 $\Omega$
Offset Delay	0.0000 ps
Length (electrical) / Offset Length	0.000 mm
Offset Loss	0.00 G $\Omega$ /s
Loss	0.0000 dB/ $\sqrt{\text{GHz}}$

**Environmental data**

Operating temperature range <sup>3</sup>	0 °C to +50 °C
Storage temperature range	- 55 °C to +90 °C
RoHS	compliant

<sup>3</sup> Temperature range over which these specifications are valid.

**Declaration of documentation**

Standard delivery for this kit includes Test Results. The documentation issued reports which quantities were tested individually, traceable to national / international standards. Model based standard definitions of the calibration standards are reported in Agilent / Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

**Inspection interval**

Recommendation	12 months
----------------	-----------

**Packing**

Weight	310 g/pce
--------	-----------

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

For the installation of the electrotechnical equipment, particular electrotechnical expertise is required.



Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
Kerstin Herzog	10.07.06	Lars Ramtke	13.06.23	j00	23-0004	Marion Striegler	25.05.23

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany <a href="http://www.rosenberger.com">www.rosenberger.com</a>				Tel. : +49 8684 18-0 Email : <a href="mailto:info@rosenberger.com">info@rosenberger.com</a>			Page 3 / 3
--	--	--	--	--	--	--	---------------