


3-1/2 digits 1999 Counts Multimeter

Item: MT-1270

- Large 3.5 digit (1999) LCD
- 0.5% best accuracy
- Data hold
- Low battery indication
- Auto power off



Digits	3 1/2 digits 1999 counts
Accuracy	0.5% best accuracy.
DCV	200mV/2V/20V/200V $\pm(0.5\%+3d)$ 1000V $\pm(0.8\%+10d)$
ACV	2V/20V/200V $\pm(0.8\%+5d)$ 750V $\pm(1.2\%+10d)$
DCA	2mA/20mA $\pm(0.8\%+10d)$, 200mA $\pm(1.2\%+8d)$, 10A $\pm(2.0\%+5d)$
ACA	20mA $\pm(1.0\%+15d)$, 200mA $\pm(2.0\%+5d)$, 10A $\pm(3.0\%+10d)$
Resistance	200 Ω $\pm(0.8\%+5d)$, 2K/20K/200K/2M Ω $\pm(0.8\%+3d)$ 20M Ω $\pm(1.0\%+25d)$, 200M Ω $\pm[5.0\%(\text{rdg}-10)+30d]$
Frequency	20K Hz $\pm(3.0\%+18d)$
Capacitance	2nF $\pm(2.5\%+20d)$, 20nF $\pm(2.5\%+20d)$, 200nF $\pm(2.5\%+20d)$, 2 μ F $\pm(2.5\%+20d)$, 20 μ F $\pm(5.0\%+30d)$
Temperature	(-20°C~400°C) $\pm 1.0\%+5d$, (400~1000°C) $\pm(1.5\%+15d)$
Battery	9V(6F22) (Not Included)
Size (HXWXD)mm	7.2 x 3.5 x 1.8 in.

Order Today! SalesSupport@EclipseTools.com

Eclipse Enterprises, Inc. ● Tel: 804.561.2610 ● Fax: 804.561.2642

13302 Chula Road, Amelia Court House, VA 23002 ● EclipseTools.com